

VIRUS TRANSPORT MEDIUM

CODE	DESCRIPTION	SPECIMEN/SAMPLE SITE
MW950S	SIGMA VIROCULT® - Standard Plastic Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Skin, nose, throat, rectum, vagina.
MW950S2	SIGMA VIROCULT® - 2 Standard Plastic Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Skin, nose, throat, rectum, vagina.
MW950SE2	SIGMA VIROCULT® - Standard Plastic + ENT Minitip Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Nasopharyngeal, urogenital
MW950S3	SIGMA VIROCULT® - 3 Standard Plastic Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Skin, nose, throat, rectum, vagina
MW950S3B	SIGMA VIROCULT® - 3 Standard Plastic Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Skin, nose, throat, rectum, vagina
MW950S3ML	SIGMA VIROCULT® - Standard Plastic Foam Printed Breakpoint - 16x100 Vial - 3ml Virocult Medium	Skin, nose, throat, rectum, vagina
MW950SENT	SIGMA VIROCULT® - ENT Minitip Foam Printed Breakpoint - 16x100 Vial - 2ml Virocult Medium	Nasopharyngeal,
MW950T	SIGMA VIROCULT® - 16x100 Vial - 2ml Virocult Medium	N/A
MW951S	SIGMA VIROCULT® - Standard Plastic Foam Printed Breakpoint - 12x80 Vial - 1ml Virocult Medium	Skin, nose, throat, rectum, vagina.
MW951S2	SIGMA VIROCULT® - 2 Standard Plastic Foam Printed Breakpoint - 12x80 Vial - 1ml Virocult Medium	Skin, nose, throat, rectum, vagina.
MW951SE2	SIGMA VIROCULT® - Standard Foam + ENT Minitip Foam Printed Breakpoint - 12x80 Vial - 1ml Virocult Medium	Nasopharyngeal, urogenital
MW951S3ML	SIGMA VIROCULT® - Standard Plastic Foam Printed Breakpoint - 12x80 Vial - 3ml Virocult Medium	Skin, nose, throat, rectum, vagina
M400001	SIGMA VIROCULT® - Ultrafine Flexible Minitip Flockswab Printed Breakpoint - 16x100 Vial - 3ml Virocult Medium	Skin, throat, nasopharyngeal
MW951S2ML	SIGMA VIROCULT® - Standard Plastic Foam Printed Breakpoint - 12x80 Vial - 2ml Virocult Medium	Skin, nose, throat, rectum, vagina
MW951SENT	SIGMA VIROCULT® - ENT Minitip Foam Swab Printed Breakpoint - 12x80Vial - 1ml Virocult Medium	Nasopharyngeal, urogenital
MW951SENT2ML	SIGMA VIROCULT® - ENT Minitip Foam Swab Printed Breakpoint - 12x80 Vial - 2ml Virocult Medium	Nasopharyngeal, urogenital
MW951T	SIGMA VIROCULT® - 12x80 Vial - 1ml Virocult Medium	N/A
MW951T3	SIGMA VIROCULT® - 12x80 Vial - 3ml Virocult Medium	N/A
MW951PF2ML	SIGMA VIROCULT® - Standard Flockswab Printed Breakpoint - 12x80 Vial - 2ml Virocult Medium	Wound, skin, throat
MW951PFENT1ML	SIGMA VIROCULT® - ENT Minitip Flockswab Printed Breakpoint - 12x80 Vial - 1ml Virocult Medium	Nasopharyngeal, urogenital
MW951SE23ML	SIGMA VIROCULT® - Standard Foam + ENT Minitip Foam Printed Breakpoint - 12x80 Vial - 3 ml Virocult Medium	Nasopharyngeal, urogenital
501355	SIGMA VIROCULT® - ENT Minitip Foam Printed Breakpoint - 12x80 RB Vial - 1ml Virocult Medium	Nasopharyngeal, urogenital
501351	SIGMA VIROCULT® - Standard Foam Printed Breakpoint - 12x80 RB Vial - 1ml Virocult Medium	Skin, nose, throat, rectum, vagina

**** Products with (Tube Only**) in the description are products which are registered as IVD's only (EU Directives and Regulations).**

VIRUS TRANSPORT MEDIUM

All other product codes are class 1s Medical Devices.

MDD: European Medical Devices Directive 93/42/EEC

IVD: European In Vitro Diagnostic Medical Devices 98/79/EC

*For Tube Only IVD's, the Notified Body number for CE, and Approved Body number for UKCA do not appear on the markings.

INTENDED USE

SIGMA VIRO CULT® Virus Collection and Transport System is intended to preserve the viability and infectivity of viral specimens for viral culture after their collection and during transport from the collection site to the testing laboratory. Σ-Virocult® specimens are processed using standard clinical laboratory procedures for viral and cell culture.

SUMMARY AND PRINCIPLES

One of the routine procedures in the diagnosis of infections caused by viruses involves the collection and transportation of a clinical swab specimen from the patient to the laboratory. Specimens containing live viruses may be submitted to a laboratory for diagnosis or confirmation of the patient's illness. Σ-Virocult® tubes contain a liquid medium to keep the specimen moist, and to maintain any viruses in a viable condition until they can be investigated at the laboratory by viral culture. The liquid medium consists of a balanced salt solution for maintaining osmotic pressure within physiological limits and phosphate buffers to stabilize the pH of the medium. For specific recommendations about the collection of specimens for viruses and primary isolation techniques, consult the following ASM publications: Cumitech 15A¹, Clinical Microbiology Procedures Handbook², Manual of Clinical Microbiology³, Clinical Virology Manual⁴, and Johnson F. B.⁵

REAGENTS

Σ-Virocult® medium is a balanced salt solution, buffered with disodium hydrogen orthophosphate, and also contains lactalbumin hydrolysate as a stabiliser, and antibiotics to inhibit the growth of any bacterial contaminants in the specimen.

Active ingredients:

Chloramphenicol

Amphotericin

Di-sodium hydrogen phosphate

Lactalbumin hydrolysate

INSTRUCTIONS FOR USE



Before use always check that immediate packaging (peel pouch) is intact, that the tube contains medium and there are no signs of leakage. In case of defect do not use the device. Appropriate protective clothing including sterile gloves should be worn when collecting and handling potentially infectious specimens. Care should be taken to avoid splashes and aerosols when snapping the swab shaft against the tube.

1. Peel back pouch, remove vial and place on a flat surface. Loosen cap.
2. Withdraw the white swab and use to take specimen.
3. Remove cap from vial, insert swab into vial and snap off the non-bud end so that the remaining shaft fits within the vial. The swabs have a scored breakpoint or moulded breakpoint to assist this process.
4. Repeat steps 2 & 3 if there is a second swab in the pack. It may have a green shaft.
5. Replace cap and turn until secure. The white swab will become attached to the cap, except for large vials and

SIGMA VIROCULT®**VIRUS TRANSPORT MEDIUM**

MW951SENT, MW951SENT2ML and MW951PFENT1ML.

6. Fill in patient's details.
7. Transport to the laboratory immediately.

PRECAUTIONS

For Professional use only.

For virology specimens only.

For in vitro Diagnostic use only.

This device is a Single use Device and therefore cannot be reused, it must be assumed that all used devices contain infections organisms and therefore should be handled accordingly, after use all devices must be disposed of according to laboratory regulations for infections waste.

DO NOT USE IF PACKAGE SEAL IS BROKEN

MATERIAL SAFETY INFORMATION

The plastic components do not contain latex or PVC.

Observe aseptic techniques and established precautions against microbiological hazards throughout all procedures. Prior to discarding, swabs and other contaminated materials must be sterilized by autoclaving. Once a swab sample is collected it should be placed immediately into the transport tube where it comes into contact with transport medium. Swab specimens for virus isolation and/or detection should be submitted to the laboratory as quickly as possible after collection.

STORAGE

Σ-Virocult® should be stored in a dry place at temperatures between + 5°C to 25°C.

DO NOT FREEZE

EXPIRY DATE

24 months from date of manufacture, expiration date is shown on the tube label, peel pouch, and box label.

SPECIMEN COLLECTION AND HANDLING

Materials provided

Swab for collection of specimen.

Transport tube with Virocult® medium

*There is no swab with MW950T, MW951T, MW951T3

Materials required but not provided

External transport container compliant with local regulations

Eagle's Minimum Essential Medium (buffered) or Hanks Balanced Salt Solution (buffered).

Pipette to withdraw 0.2ml fluid from Virocult® tube.

Cell culture facilities and cell lines appropriate for target viruses.

PROCESSING SPECIMENS

Refer to the test system manufacturer's instructions. Any use with non-culture methods must be validated by the user.

QUALITY CONTROL

With reference to CLSI M40-A it is recommended that Herpes Simplex Type 2 ATCC VR-734 be used as a control strain. The swab is inoculated from a suspension containing 5×10^4 TCID per ml and placed into the transport tube. The tube is held at the desired transport temperature (4°C or room temperature) for up to 96 hours. The transport tube is processed as described above ("Processing Specimens") and 0.2ml of suspension is inoculated onto a suitable tissue culture monolayer. Any recovery of virus is acceptable performance.

VIRUS TRANSPORT MEDIUM

LIMITATIONS

1. This device is NOT SUITABLE FOR THE TRANSPORT OF BACTERIA OR FUNGI because antibiotics are used in the medium.
2. Traces of antiviral reagents at site of specimen prior to sampling may affect recovery of virus.
3. Σ-Virocult® has not been validated for use with molecular techniques such as PCR, so any such use must be validated by the user.

REFERENCES

1. Gleaves C. A., R. L. Hodinka, S. L. G. Johnston and E. M. Swierkosz, Cumitech 15A. Laboratory Diagnosis of Viral Infections, p. 7., American Society for Microbiology, Washington D.C., 1994
2. Miller, M. J., and A.L. Warford. Preparation of specimens for inoculation of cell cultures, p. 831-838 In H.D. Isenberg (ed.), Clinical Microbiology Procedures Handbook. American Society for Microbiology, Washington, D.C., 1992.
3. Chapin, K.C., & F.W. Westenfeld, 2003, Reagents, Stains, Media, and Cell Lines: Virology, p.1250 in Murray P.R., E.J. Baron, J.H. Jorgensen, M.A. Pfaller, & R.H. Tenenbaum, 2003, Manual of Clinical Microbiology, 8th Edition, ASM Press, Washington D.C.
4. Specter, S., R.L. Hodinka, and S.A. Young, 2000, Clinical Virology Manual, 3rd Edition, ASM Press, Washington D.C.
5. Johnson F. B., Transport of Viral Specimens, p. 120 – 131. Clinical Microbiology Reviews, Vol. 3, No. 2, April 1990
6. CLSI. 'Quality Control of Microbiological Transport Systems'; Approved Standard M40-A. CLSI (formerly NCCLS) document M40-A [ISBN 1-56238-520-8]. CLSI, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2003.

* HydraFlock® and PurFlock® are registered trademarks of Puritan Medical Products LLC

** Products with (Tube Only*) in the description are products which are registered as IVD's only (EU Directives and Regulations) .

REGULATORY SYMBOLS APPLICABLE FOR FAMILY GROUP



EC REP Advena Ltd, Tower Business Centre, 2nd Fl.,
Tower Street, Swatar, BKR 4013, Malta

CH REP Swiss AR Services AG Industriestrasse 47
CH-6300 / Zug

SYMBOLS & DEFINITIONS



SIGMA VIROCULT®


VIRUS TRANSPORT MEDIUM



Use By (YYYY/MM)



Consult Instructions for Use



Date of Manufacture



Peel Here



UKCA Mark



Contains Sufficient for <n> Tests

REF

Catalogue Number

CH

REP

Swiss Representative

LOT

Batch Code

